DEPARTMENT OF CIVIL ENGINEERING
ELLIS HALL

Queens University
Kingston, Canada
K7L 3N6
Tel 613 545-2122
Fax 613 545-2128

February 8, 1993

PRIVILEGED & CONFIDENTIAL
ATTORNEY WORK PRODUCT
Do Not Distribute or Reproduce

Jeffrey Lawson
Environmental Project Control
Two Grafton Common
Post Office Box 536
Grafton, MA
U.S.A. 01519

Dear Jeffrey:

At the request of Jeffrey Bates of Goodwin, Proctor & Hoar, I have completed both interfacial tension and surface tension measurements on the sample of DNAPL (sample 172119) shipped to me January 26, 1993. These measurements were performed in triplicate at 20 °C using a KRUSS Model K8 platinum-iridium ring interfacial tensiometer. Details of the results are as follows:

1)	Water-DNAPL interfacial tension:	16.7 dynes/cm
		16.8 dynes/cm
		16.8 dynes/cm
	Average of 3 readings:	**16.77 dynes/cm**
2)	DNAPL-air surface tension:	34.8 dynes/cm
		34.3 dynes/cm
		34.5 dynes/cm
	Average of 3 readings:	**34.53 dynes/cm**

If you have any questions regarding these measurements, please do not hesitate to call me at (613) 545-6834 at your earliest convenience.

Sincerely yours,

Bernard H. Kueper, Ph.D.
Assistant Professor